

Poster Sponsor:  
**Baxter**  
Baxter Transfusion Therapies  
One Baxter Parkway  
Deerfield, IL 60015  
USA  
847.948.2000  
www.baxter.com

Poster Sponsor:  
**CERUS**  
Cerus Corporation  
2411 Stanwell Drive  
Concord, CA 94520  
USA  
1.925.288.6000  
www.cerus.com

Primary Author:  
**Ao. Univ. Prof. Dr. C. Lass-Flörl**  
Medizinische Universität Innsbruck  
Krankenhaushygiene  
Fritz Pregl Str. 3/III  
6020 Innsbruck  
0043 512 507 3425  
cornelia.lass-flörl@uibk.ac.at



## **Comparative Efficacy of Bacterial Culture Versus Pathogen Inactivation to Reduce the Risk of Transfusion-associated Sepsis**

**C. Lass-Flörl MD<sup>1</sup>, D. Allersdorfer<sup>1</sup>, M. Rheinschmidt<sup>2</sup>, L. Lin PhD<sup>2</sup>,  
D. Schönitzer MD<sup>3</sup> and W. Nussbaumer MD<sup>3</sup>**

1. Institute for Hygiene and Social Medicine, University of Innsbruck, Innsbruck, Austria;
2. Cerus Corporation, Concord, CA, USA;
3. Department of Transfusion Medicine, University of Innsbruck, Innsbruck, Austria

**Presented at the  
XXII Congr s de la Soci t  Fran aise  
de Transfusion Sanguine (SFTS)  
Saint-Malo, France • June 27 & 28, 2005**

## Objective


This study compared the efficacy of bacterial screening using a culture method (BacT/Alert System, Bioré) with pathogen inactivation (INTERCEPT Blood System, Baxter Healthcare) to prevent the transfusion of platelet components contaminated with bacteria.

## Methods

Seven species of bacteria associated with transfusion-associated sepsis, including Gram-positive aerobes and anaerobes and Gram-negative aerobes were studied. Paired (Control and Test) platelet concentrates (PC) containing approximately  $3 \times 10^{11}$  platelets in 300 mL of 35% plasma and 65% InterSol<sup>®</sup> from double-dose apheresis collections were contaminated with each species of bacteria at approximately 10, 100, and

1,000 cfu per unit. After an overnight storage, the Test PC was treated with INTERCEPT (150  $\mu$ M amotosalen and 3 J/cm<sup>2</sup> UVA, using steps 1 to 3 in Figure 1). The Control PC was not treated. Both units were assayed for aerobic and anaerobic bacteria using the BacT/Alert System on days 1, 2 and 5 of storage. A platelet sample was considered contaminated with bacteria if a positive signal was registered within 120 hrs of culture.

Figure 1: INTERCEPT System for Platelets



## Results

Cultures failed to detect some low-dose inocula in Control PC (Table 1 and Figures 2 to 4). All Test

PC treated with INTERCEPT remained negative by BacT/Alert cultures throughout the entire 5-day

observation period regardless of the bacterial species and the contamination level (Table 1 and Figure 5).

Figure 2, Bacterial detection: results from untreated platelet concentrates (Day-1 samples)




Figure 3, Bacterial detection: results from untreated platelet concentrates (Day-2 samples)




Figure 4, Bacterial detection: results from untreated platelet concentrates (Day-5 samples)




Figure 5: INTERCEPT treated platelet concentrates remained all negative




Table 1:

	Time (h=hours) to Positive Culture*					
	10 cfu/unit		100 cfu/unit		1,000 cfu/unit	
	aerobe	anaerobe	aerobe	anaerobe	aerobe	anaerobe
<b>Inoculated Control PC (Untreated)</b>						
<i>Staphylococcus aureus</i>	14 h	16 h	11 h	12 h	7 h	7 h
<i>Staphylococcus epidermidis</i>	Neg	D5 +30 h	50 h	35 h	16 h	16 h
<i>Propionibacterium acnes</i>	Neg	Neg	Neg	D5 +50 h	Neg	D5 +50h
<i>Escherichia coli</i>	Neg	Neg	Neg	Neg	4 h	4 h
<i>Clostridium perfringens</i>	Neg	Neg	Neg	Neg	11 h	D2 +10 h
<i>Streptococcus agalactiae</i>	Neg	Neg	D2 +8 h	13 h	8 h	8 h
<i>Klebsiella pneumonia</i>	D2 +8 h	D2 +8 h	D2 +7 h	D2 +7 h	8 h	8 h
<b>Inoculated Test PC (treated with INTERCEPT)</b>						
<i>Staphylococcus aureus</i>	Neg	Neg	Neg	Neg	Neg	Neg
<i>Staphylococcus epidermidis</i>	Neg	Neg	Neg	Neg	Neg	Neg
<i>Propionibacterium acnes</i>	Neg	Neg	Neg	Neg	Neg	Neg
<i>Escherichia coli</i>	Neg	Neg	Neg	Neg	Neg	Neg
<i>Clostridium perfringens</i>	Neg	Neg	Neg	Neg	Neg	Neg
<i>Streptococcus agalactiae</i>	Neg	Neg	Neg	Neg	Neg	Neg
<i>Klebsiella pneumonia</i>	Neg	Neg	Neg	Neg	Neg	Neg

\* D = day sampled, where not indicated, samples were withdrawn on day 1.  
Neg = No bacteria detected even when sampled on day 5 of platelet storage and after 120 h of culture.

## Conclusions

- Bacterial detection using cultures may fail to detect low levels of bacteria.
- Failure to detect bacteria will result in the release of contaminated platelet products with “test negative-to-date” status.
- Inactivation of bacteria using the INTERCEPT Blood System is capable of preventing release of contaminated platelet components.